


CONTROLLED
2/23/05

REV	DESCRIPTION OF CHANGE	CHANGE REFERENCE	ORIGINATOR	VERIFIED	APPROVED
			DATE AND SIGN	DATE AND SIGN	DATE AND SIGN
			DATE AND SIGN	DATE AND SIGN	DATE AND SIGN
A	Release	N/A	DATE AND SIGN MY 2/23/05	DATE AND SIGN 2/22/05	DATE AND SIGN 2/23/05
THE INFORMATION CONTAINED HEREIN IS PROPRIETARY TO LAERDAL AND SHALL NOT BE USED FOR ANY PURPOSE DETRIMENTAL TO OR BE REPRODUCED WITHOUT PERMISSION OF LAERDAL		DMR <input type="checkbox"/> DHF <input type="checkbox"/>	 Laerdal helping save lives		
REFERENCE	N/A	REPLACES	N/A	REV	N/A
TITLE					
DFU Extri Kelly					
PRODUCT GROUP	TOOL	PART NO	REV	DOCUMENT NO	PAGE
EX-KELLY	N/A	1001604	A	PRO-ML03-0022	1 of 1
				ATTACH	REV
				1	A

EXTRI KELLY

Directions for Use



Laerdal[™]
helping save lives

Table of Contents

Laerdal Recommends	4
Items Included	5
Skill Taught	5
Head	5
Chest Plate	5
Belly Plate	5
Waist, Elbow, Knee & Ankle Joints	5
Arms	6
Legs	6
Care and Maintenance	7
Warranty	8
Replacement Parts	9

Extri Kelly is a full-body, lifelike manikin which realistically simulates a patient. It is specifically designed for training professionals in the practice of extrication.

As a result of quality construction and the ease of replacing individual parts, this simulator should provide many sessions of training when reasonable care and maintenance are practiced.

Laerdal Recommends:

When assembled, talcum powder should be used in joints to assist articulation.

Instructor Notes:

Item Included:

(1) Extri Kelly Manikin

Skill Taught:

- Extrication

Head: Head is connected to torso by a long bolt at base of neck, secured with a wing nut. Remove chest plate from torso and unscrew wing nut. Head will slip out of torso.

Chest Plate: Chest plate fits into torso by a ledge on plate and corresponding groove on torso cavity. Remove plate by applying pressure to center of plate and working your fingers under edge of plate and lifting out. To insert, position plate over cavity and work edges into groove on torso cavity.

Belly Plate: Belly plate fits into lower torso by a ledge on plate and corresponding groove on abdominal cavity. Remove plate by applying pressure to center of plate and working your fingers under edge of plate and lifting out. To insert, position plate over cavity and work edges into groove on abdominal cavity.

Laerdal Recommends

When assembled, talcum powder should be used in joints to assist articulation.

Waist, Elbow, Knee and Ankle Joints:

A bolt and lock nut connect these joints.

To remove:

1. Unscrew lock nut.
2. Pull bolt out of joint.

The segments will now easily disconnect.

Laerdal Recommends

When assembled, talcum powder should be used in joints to assist articulation.

Arms: Arms are attached using a 4 ½ inches long bolt, three washers, a spring and a wing nut.

Laerdal Recommends


Attachment hardware is located in envelope in box.

To attach:

1. Remove chest plate from manikin to locate hole in shoulder.
2. Remove deltoid injection pad from arm by squeezing skin in center of pad while pulling
3. Slide washer over bolt.
4. Insert bolt through hole in arm and then through hole in the manikin's shoulder.
5. Slide one washer, a spring, a second washer and a wing nut over bolt.
6. Use screwdriver to hold bolt steady while tightening wing nut. (See Fig. 1)

To remove arms, reverse procedure.

Fig. 1




Legs: Legs are attached using a 6 inches long bolt, a spring, a large washer, a small washer and a wing nut.

To attach:

1. Remove belly plate from manikin to locate hole in hip.
2. Remove thigh injection pad from leg by squeezing skin in center of pad while pulling.
3. Slide spring and small washer over bolt.
4. Insert bolt through hole in pelvis.
5. Slide large washer and wing nut over bolt.
6. Use a screwdriver to drive bolt into wing nut. (Fig. 2)

To remove legs, reverse procedure.

Fig. 2



Care and Maintenance:

1. Clean with mild soap and water; do not submerge manikin or parts in cleaning fluids or water.
2. Use only on clean surface. Avoid felt tipped markers, ink pens, acetone, iodine or other staining products and avoid placing the manikin on newsprint or inked lines of any kind.
3. To ensure longevity, each manikin and module should be cleaned after each training session and a general inspection should be conducted regularly.
4. Modules and all other parts should be drained and air-dried thoroughly before storage and disinfected when needed. After use of injection pads (*use water only*), accumulated water should be squeezed out. **Do not store wet foam pads in the skin.** To prevent mildew or mold, pads can be soaked in a mild solution of disinfectant and water or bleach and water. Squeeze excess solution from pads, allow them to dry, then store or reinsert in manikin.
5. Articulating parts will benefit from a light application of talcum powder prior to training sessions.
6. Store properly between teaching sessions.

Warranty

Laerdal Medical warrants to the purchaser that its products are free from defects in material and workmanship for a period of one (1) year from the date of purchase by the original user. During the designated one (1) year period, Laerdal Medical will, upon receipt of a product found to be defective due to materials or workmanship from the purchaser and notification in writing of the defect, at its option repair or replace any parts found to be defective or the entire product. Warranty period does not renew with replacement or repair of original product.

Products found to be defective and notification of defects may also be sent to the authorized Laerdal Medical dealer from whom the product was purchased. All postage, shipping or handling charges shall be the sole responsibility of the purchaser.

Laerdal Medical is responsible for the effects of safety, reliability and performance of its product(s) only if:

- service, repair, readjustment or modification is carried out by Laerdal Medical or persons authorized by Laerdal Medical.
- the electrical installation of the room where the product is used complies with pertinent equipment requirements.
- the product is used in the proper manner in strict compliance with its Directions for Use.

Laerdal Medical shall not be liable under this warranty for incidental or consequential damages, or in the event any unauthorized repairs or modifications have been made or attempted, or when the product, or any part thereof, has been damaged by accident, misuse or abuse. This warranty does not cover batteries, fuses, normal wear and tear, excessive use, staining, discoloration or other cosmetic irregularity which does not impede or degrade product performance.

Some states in the USA do not allow the exclusion or limitation of incidental or consequential damages, so those limitations or exclusions may not apply to you.

There are no other express or implied warranties, whether of merchantability, fitness or purpose, or otherwise, on the product, its parts and accessories.

Replacement Parts

<u>Product Number:</u>	<u>Description:</u>
380460	Genitalia, Blank-Adult Male
380471	Plug, Belly plate-Adult
380475	Pin, Pelvis
380600	Leg Assembly, Left Adult
380650	Leg Assembly, Right Adult
380800	Arm Assembly, Left Plain Adult
380801	Hand, Left Adult Male
381401	Arm, Upper Right Adult Male
381402	Pad Assembly, Deltoid Injection Adult
101-10150	Belly plate, Plain Hospital Adult
101-10250	Torso, Hospital Half Cut Adult
101-10350	Pelvis, Hospital W/Bridge Adult
101-10450	Hardware Set, Extri Kelly Arms/Legs/Tools
300-00150	Pad Assembly, Thigh Injection Adult
300-00250	Pad Assembly, Ventral Gluteal Injection Adult Male
300-00350	Pad Assembly, Gluteal Injection Adult Male
300-02350	Forearm, Right Adult Male
300-02450	Pad, Forearm Right Non-Injection Adult Male
300-02550	Hand, Right Adult Male
300-02650	Arm, Upper Left Adult Male
300-02750	Forearm, Left Adult Male
300-02850	Pad, Forearm Left Non-Injection Adult Male
300-02950	Thigh, Right Adult
300-03050	Leg, Lower Right Adult

300-03150	Foot, Right Adult
300-03250	Thigh, Left Adult
300-03350	Leg, Lower Left Adult
300-03450	Foot, Left Adult
300-04850	Gown, Hospital-Back Tie
300-05150	Arm Assembly, Left Plain Adult
300-10250	Chest plate, Hospital Plain Adult Male
301-10150	Head Assembly, Hard Adult Male

Please contact Customer Service Representatives for more information on Replacement Parts in other skin tones.

(C) Copyright Medical Plastics Laboratory, Inc., Gatesville TX
Document #I001604 rev A
February 5, 2003

Norway: LAERDAL MEDICAL AS
P.O. Box 377, N-4002 Stavanger
Tel. +47 51 51 17 00, Fax +47 51 52 35 57
E-mail: laerdal.norway@laerdal.no

Australia: LAERDAL Pty. Ltd. (ABN 47 003 817 490)
23 Edward Street, Oakleigh, Victoria 3166
Tel. +61 3 9569 4055, Toll free 1800 331 565
Fax +61 3 9569 4028, Toll free 1800 635 835
E-mail: customer.service@laerdal.com.au

Benelux: "The Laerdal BeNeLux House"
Dijkstraat 54, NL-5554 PS VALKENSWAARD
Tel. +31 40 208 58 00, Fax +31 40 208 58 02
E-mail: laerdal.benelux@laerdal.no

Canada: LAERDAL MEDICAL CANADA LTD.
151 Nashdene Rd., Unit #45
Toronto, ON, Canada, M1V 4C3
Tel. +1 (416) 298-9600,
Toll free 888/LAERDAL (523-7325)
ou en français (800) 567-9987
Fax +1 (416) 298-8016
E-mail: savelives@laerdal.ca

Denmark: LAERDAL DANMARK
Postbox 225, DK-8600 Silkeborg
Tel. +45 80 333 112, Fax +45 80 333 555
E-mail: laerdal.denmark@laerdal.no

Eastern Europe: LAERDAL C.E. EUROPE
A-1060 Vienna, Mariahilferstr. 1d/1/3
Tel - Fax +43 15877140, Tel. +43 15815927
E-mail: mstam@aon.at

Far East: LAERDAL Singapore Pte Ltd
No 1 Marine Parade Central
#13-05, Parkway Centre Singapore 449408
Tel. +65 63464259, Fax +65 63467523
E-mail: admin@laerdal.com.sg

Finland: LAERDAL OY
Vattuniemenranta 2, 00210 HELSINKI
Puhelin +358 (0)9-612 99 80,
Telekopio +358 (0)9-692 77 99
E-mail: laerdal.finland@laerdal.no

France: LAERDAL MEDICAL FRANCE
1 rue des Vergers - Bâtiment n° 5 - 69578 Limonest Cedex
Tél. +33 (0)4 72 52 02 52, Fax +33 (0)4 78 35 38 45
E-mail: laerdal.france@laerdal.no

Germany: LAERDAL MEDICAL DEUTCHLAND GmbH
Am Loferfeld 56, 81249 München
Tel. +49 (0)89 / 864 954-0, Fax +49 (0)89 / 864 34 84
E-mail: info@laerdal.de

Italy: LAERDAL ITALIA s.r.l.
Via Piero Gobetti 52/2 int.Z, 40129 Bologna
Tel. +39 051-355587, Fax +39 051-355598
E-mail: laerdal.italia@laerdal.no.

Japan: LAERDAL MEDICAL JAPAN K.K..
Shinjuku IS Bldg. 8F
2-13-12, Shinjuku, Shinjuku-ku, Tokyo 160-0022
Tel. 81-3-5919-2121, Fax: 81-3-5919-2122
E-mail: laerdal.japan@laerdal.co.jp

Malaysia: LAERDAL HOSPILINE Sdn Bhd
12 Jalan Titir 33/25, Section 33, Shah Alam Technology Park,
40400 Shah Alam, Selangor Darul Ehsan,
Tel +60 (3) 5122-7002, Fax +60 (3) 5122-5650
E-mail: sales@hospiline.com

New Zealand: LAERDAL NEW ZEALAND Ltd
(GST Number 79-403-636)
PO Box 302664, North Harbour, Auckland 1330
Tel. +61 3 9564 4066, Toll free 0800 523 732
Fax +61 3 9563 3368, Toll free 0800 528 852
E-mail: customer.service@laerdal.co.nz

Portugal: LAERDAL ESPANA, Sucursal em Portugal
Rua da Bela Vista à Graca, no.31 A. Escritorio 9,
1170-054 Lisboa
Tel. (+351)-21-816-6567, Fax (+351)-21-816-6568,
Mob. (+351)-918765367

Spain: LAERDAL ESPANA, S.L.
Manuel Tovar 19, 28034 Madrid
Tel. (+34) 902-291110, Fax (+34) (91)-7291125
E-mail: laerdal.spain@laerdal.no

Sweden: LAERDAL MEDICAL AB
Box 2064, Paviljongvägen 3-5, 132 02 Saltsjö-Boo
Tel. +46 (0)8-55614610, Fax +46 (0)8-55614619
E-mail: Laerdal.sweden@laerdal.se

United Kingdom: LAERDAL MEDICAL LTD.
Laerdal House, Goodmead Road, Orpington,
Kent BR6 0HX
Tel. +44 (0)1689 876634, Fax +44 (0)1689 873800
E-mail: customer.service@laerdal.co.uk

USA and Latin America:
LAERDAL MEDICAL CORPORATION
167 Myers Corners Road, P.O. Box 1840
Wappingers Falls, New York 12590-8840
Tel. (800) 431-1055, +1 (845) 297-7770
Fax (800) 227-1143, +1 (845) 298-4545
E-mail: customerservice@laerdal.com

MEDICAL PLASTICS LABORATORY, INC.
P.O. Box 38, 226 FM 116 South
Gatesville, Texas 76528-0038
Tel. (800) 433-5539, (254) 865-7221
Telefax (254) 865-8011
E-mail: services@mpltx.com